



AB 994

*Research laboratory  
accredited by PCA, Nr AB  
994*

*Scopes of accreditation:*

- concentration and mass measurements of flow of fine particles
- measurements of concentration and mass flow of SO<sub>2</sub>, NO<sub>x</sub>, CO
- concentration measurements of CO<sub>2</sub>, O<sub>2</sub>
- concentration and mass flow measurements of O<sub>2</sub>O
- sampling for mass concentration determination of PCDD/PCDF and dioxin type PCB
- sampling for concentration determination of (As; Cd; Cr; Co; Cu; Mn; Ni; Pb; Sb; Tl; V)
- sampling for concentration determination of Hg
- sampling and determination of concentration and mass flow of HCl
- sampling and determination of concentration and mass flow of HF
- sampling for determining the concentration of individual gaseous organic compounds
- calibration of Automated Monitoring systems,
- QAL2 procedure
- annual performance test of Automated Monitoring Systems, AST procedure
- noise measurement from machinery, installations and industrial plants

Chorzów, November 16<sup>th</sup> 2023  
Our ref. No.: PW/25/11/23

## **Report No PW/25/11/23**

**on concentration measurements of PCDDs/PCDFs and dioxins-like PCBs emitted into environment from flue gas channel located at UAB GREN KLAIPEDA, Kretainio g. 3 LT-94103, Klaipeda, Lithuania.**

Client name and address:

**UAB "GREN KLAIPEDA"**

Kretainio g. 3,  
LT-94103 Klaipeda

Developed by:

mgr Grzegorz Bortel

inż. Dariusz Guja

inż. Dariusz Guja

DYREKTOR

inż. Dariusz Guja

Calculation and data transfer checked by:

Verified by:

Copy number: 1/2

---

**Test Report No PW/25/11/23**

---

**List of contents:**

- 1. PURPOSE AND SCOPE OF THE REPORT**
- 2. BASIS OF MEASUREMENTS EXECUTION**
- 3. MEASUREMENT TEAM**
- 4. MEASUREMENT RESULTS SUMMARY**
- 5. DESCRIPTION OF THE MEASUREMENT METHOD**
- 6. MEASUREMENT RESULTS**
- 7. MEASUREMENT DEVICES**
- 8. ANALYSIS RESULTS**
- 9. CERTIFICATE OF ACCREDITATION**
- 10. MEASUREMENT PLANE SCHEME**

**Test Report No PW/25/11/23****1. PURPOSE AND SCOPE OF THE REPORT**

Purpose of executed measurements was to determine the concentration and emission of *PCDDs/PCDFs and dioxins-like PCBs* emitted to environment from flue gas channel located at UAB GREN KLAIPEDA, Kretainio g. 3 LT-94103, Klaipeda, Lithuania

*Measurements range:*

- PCDD/DF emission and concentration,
- PCB emission and concentration.

The measurements were carried out in accordance with the sampling plan and the described sampling methods.

Operating parameters of the technological installation was obtained from the customer's representative.

**2. BASIS OF MEASUREMENTS EXECUTION**

The measurements were taken according to the Purchase Order No MX27936LTKLJ21 dated May 07<sup>th</sup> 2021, our reference number PP/01/05/21.

**3. MEASUREMENT TEAM**

The measurements taken on October 25<sup>th</sup> 2023 were executed by the following team:

- Grzegorz Bortel specialist- measurement team leader,
- Bartłomiej Glik specialist,
- Grzegorz Kurzeja specialist.

**4. MEASUREMENT RESULTS SUMMARY**

Below are presented measurement results summary, full measurement results are presented in chapter no 6, at page 8.

Concentration of the substance in the gas in the reference conditions O2 ref. 11%	PCDDF	ng/m <sup>3</sup>	0,0023
	PCB	ng/m <sup>3</sup>	0,0008
Emission limits	PCDDF	ng/m <sup>3</sup>	0,1

---

**Test Report No PW/25/11/23**

---

**5. DESCRIPTION OF THE MEASUREMENT METHOD****Measurement of the gas volumetric flow**

The flow rate was determined according to ISO Standard PN-EN ISO 16911-1:2013 „Stationary source emissions - Manual and automatic determination of velocity and volume flow rate in ducts - Part 1: Manual reference method”. Gravimetric dust monitor type Megasystem APIS X-1 and type "S" Pitot tube were used for the measurements. Measurement is accredited.

Accreditation range: differential pressure: > 5 Pa

**Measurement O<sub>2</sub> content**

The concentration of O<sub>2</sub> was determined using gas analyzer HORIBA PG-350E-EU equipped with testing probe 1750 mm long. The measurements were taken according to the procedure described in measurement unit as well as to Polish Standard PN-EN 14789:2017 "Stationary source emissions - Determination of volume concentration of oxygen O<sub>2</sub> - Reference method - Paramagnetism". Measurement is accredited.

Accreditation range: O<sub>2</sub> content: 3-21%

**Measurement CO<sub>2</sub> content**

The concentration of CO<sub>2</sub> was determined using gas analyzer HORIBA PG-350E-EU equipped with testing probe 1750 mm long. The measurements were taken according to the procedure described in measurement unit as well as to Polish Standard PN-ISO 10396:2001 "Stationary Source Emissions - Sampling For The Automated Determination Of Gas Concentrations ". Measurement is accredited.

Accreditation range: CO<sub>2</sub> content: 0,1-20%

**Measurement moisture content**

Moisture content was determined using condensation-absorption method. The measurements were taken according to Polish Standard PN-EN 14790:2017. Measurement is accredited.

Accreditation range: H<sub>2</sub>O content: 29-250 g/m<sup>3</sup>

---

**Test Report No PW/25/11/23**

---

**PCDDs/PCDFs and dioxins-like PCBs sampling and determination**

PCDD/DF samples were taken according to the requirements of Polish Standard PN-EN 1948-1:2006 „ Stationary source emissions - Determination of mass concentration of PCDDs/PCDFs and dioxin-type PCBs - Part 1: Sampling of PCDDs/PCDFs”.

PCB samples were taken according to the requirements of Polish Standard PN-EN 1948-4:2014-03 „ Stationary source emissions - Determination of mass concentration of PCDDs/PCDFs and dioxin-type PCBs - Part 4: Sampling and analysis dioxin-like PCBs”.

*The three stages of PCDD/DF and PCBs concentration and emission determination:*

**Stage I - sampling**

For the determination of mass concentration of PCDD/DF and PCBs proper sampling plays important role that affects following stages of the testing. The sampling were performed by means of the filtration and condensation method using PCDD/DF and PCBs sampling system conformed to European Standard PN-EN 1948-1:2006 and PN-EN 1948-4:2014-03.

The following page shows the schematic diagram of the sampling system.


**Stage II - laboratory analysis**

The samples were analysed at the ALS Czech Republic, s.r.o accredited laboratory following the CSN EN 1948-2,3,4 methodology: The determination of polychlorinated dibenzo-p-dioxine, dibenzofurans in emission samples with the method of isotop dillution using HRGC/HRMS.

The dioxin and furans analyses were conducted at the ALS Czech Republic, s.r.o. Laboratory, specifically accredited for the test by the Český Institut Pro Akreditaci, o.p.s. , Nr L 1163.

**Stage III - development and the results and discussion**

The last stage includes results collection from the previous stages, emission calculation and PCDD/DF and PCBs concentration as well as comparison to the standards in force.

**Test Report No PW/25/11/23**

Opis:

- 1. ferrule
- 2. cooler
- 3. condensation pot
- 4,18. aspiration path - elbows
- 5,10. aspiration path - tubes
- 6-8. scrubbers
- 9. aspiration path - probe tube
- 11,12,14,15. – aspiration endings
- 13. clamp
- 16. holding springs
- 17. aspiration controller
- 19. temperature controller
- 20. Pitot tube „S” - type
- 21. filter casing
- 22. yoke
- 24. gas dryer
- 25. cooler casing
- 26. heated probe
- 28. cooler coil
- 29,30. support

---

**Test Report No PW/25/11/23**

---

## **6. MEASUREMENT RESULTS**

**Test Report No PW/25/11/23**

- 1) Plant or unit name: Combined Heat and Power Plant<sup>1)</sup>  
 2) Flue gases cleaning unit: ALSTOM NID flue gas cleaning solution (bag filters, activate carbon, ammonia solution, hydrated lime)<sup>1)</sup>  
 3) Emission source load during measurements: ~95,33 MW<sup>1)</sup>  
 4) Fuel type or material mass flow in process: 45% municipal waste, 55% industrial waste: 34,3 t/h<sup>1)</sup>  
 5) Location of sampling and measurements: in duct, after flue gases cleaning unit

Measurement reference number		25-10-01					
Date of measurement		25.10.2023					
Measurement time range		09:56-15:57					
Scope of test		Unit	Results	Uncertainty +/-	Method		
Meteorological conditions	Atmospheric pressure	hPa	1003,1		PN-Z-04030-7:1994		
	Air temperature	oC	11				
Measurement plane	Diameter	m	2,20				
	Area	m <sup>2</sup>	3,7994				
Stack gas parameters	Temperature	oC	40		PN-Z-04030-7:1994		
	Static pressure	Pa	-7				
	Differential pressure	Pa	146				
	Gas moistness grade X	kg/kg	0,060				
	Average velocity	m/s	12,3	0,8			
	Chemical composition	O <sub>2</sub>	%	7,7	PN-EN 14789:2017		
		CO <sub>2</sub>	%	11,6	PN-ISO 10396:2001		
	Wet gas density during testing	kg/m <sup>3</sup>	1,119		PN-EN 14790:2017		
	Gas density in normal conditions	kg/m <sup>3</sup> N	1,297		PN-EN 14790:2017		
	Gas density in conventional conditions	kg/m <sup>3</sup> U	1,347		PN-EN 14790:2017		
Concentration in the gas at measurement conditions	PCDDF*	ng/m <sup>3</sup>	0,0025	0,0008	PN-EN 1948:2006		
	PCB*		0,0009	0,0003			
Concentration in the gas at normal conditions	PCDDF*	ng/m <sup>3</sup> N	0,0029	0,0009	PN-EN 1948:2006		
	PCB*		0,0011	0,0003	PN-EN 1948-4:2014-03		
Concentration of the substance in the gas in the standard conditions	PCDDF*	ng/m <sup>3</sup> U	0,0030	0,0009	PN-EN 1948:2006		
	PCB*		0,0011	0,0003	PN-EN 1948-4:2014-03		
Concentration of the substance in the gas in the reference conditions O <sub>2</sub> ref. 11%	PCDDF*	ng/m <sup>3</sup> U	0,0023	0,0007	PN-EN 1948:2006		
	PCB*		0,0008	0,0003	PN-EN 1948-4:2014-03		
Gas volume flow	measurement conditions	m <sup>3</sup> /h	167554	21034	PN-EN ISO 16911-1:2013		
	normal conditions	m <sup>3</sup> N/h	144610	18154			
	conventional conditions	m <sup>3</sup> U/h	131343	16489			
	conventional conditions O <sub>2</sub> ref. 11%	m <sup>3</sup> U/h	174686	22304			
The emission obtained by measuring	PCDDF*	ng/h	394,03	132,62	PN-EN 1948:2006		
	PCB*		144,48	48,63	PN-EN 1948-4:2014-03		
Emission limits	PCDDF*	ng/m <sup>3</sup> U	0,1				
Transgession	PCDDF*	ng/m <sup>3</sup> U	-				

\*- the results obtained from the subcontractor ( accredited )

<sup>1)</sup>-information obtained from the client

**Test Report No PW/25/11/23**Notes:

Normal conditions designate the temperature of 273 K and pressure of 101,3kPa, defining normal cubic meter m<sup>3</sup>N. The conventional conditions designate the temperature of 273K, pressure of 101,3 kPa and dry gases (steam contents less than 5 g/kg of flue gas), defining conventional cubic meter, m<sup>3</sup>U

The specified expanded uncertainty comes from standard uncertainty multiplied by expansion coefficient k = 2, which provides 95% level of confidence for normal distribution. Uncertainty takes into account the sampling and analysis.

Registry of samples delivered to the laboratory: P/38/10/23, P/39/10/23

Date of delivery to the laboratory: 27.10.2023

Date of analysis: 27.10 - 14.11.2023

**Field blanks:**

ID/ number of sample	Type of substance	The criterion of the blank [ng/m <sup>3</sup> ] 11%O <sub>2</sub>	The value of the blank [ng/m <sup>3</sup> ] 11% O <sub>2</sub>	Result [+/-]
P/39/10/23	PCDD/DF	0,01	< 0,0020	+
	PCB	n/a	< 0,0008	+

Work parameters of measurement system:**PCDD/DF (PN-EN 1948:2006) and PCB (PN-EN 1948-4+A1:2013)::**

sampling method:	condensation - adsorption method		
filter paramethers:	19 x 90 mm, filter efficiency: 99,998 %		
sampling train:	2 measurement axis		
oxygen reference :	11 %		
time of dioxins and furans measerument:	09:56 – 15:57 (360,1 min)		
nozzle diameter:	6	mm	
probe temperature:	119	°C	
scrubbers temperature	5	°C	
aspired gas volume	5,915	m <sup>3</sup>	
average sampling flow	16,4	l/min	
izokinetic ratio:	95,4	%	
leak test:	+ / +		
gas meter temperature	10,5	°C	
gas meter pressure	0	bar	
spiking pattern:	filter surface		
absorption solution:	100 ml H <sub>2</sub> O dest. + 50 ml 2-etoksyetanol		
recovery:	90 %	<sup>13</sup> C <sub>12</sub> -1,2,3,7,8-PECDF,	/>50%/
	84 %	<sup>13</sup> C <sub>12</sub> -1,2,3,7,8,9-HxCDF,	/>50%/
	90 %	<sup>13</sup> C <sub>12</sub> -1,2,3,7,8,9 HpCDF.	/>50%/
	70 %	<sup>13</sup> C <sub>12</sub> -2,3,4,4'-tetraCB (60),	/>50%/
	93 %	<sup>13</sup> C <sub>12</sub> -3,3',4,5,5'-pentaCB (127),	/>50%/
	73 %	<sup>13</sup> C <sub>12</sub> -2,3,3',4,5,5'-hexaCB (159).	/>50%/
PCDD/DF TEQ sample mass:	0,017	ng	
PCB TEQ sample mass:	0,0061	ng	

**Test Report No PW/25/11/23****H<sub>2</sub>O (PN-EN 14790:2017)**

sampling train: 2 measurement axis  
 sampling equipment: titanium sampling train  
 heated probe 2,0 m long  
 sampling pump: SKC  
 cartridge: moisture measurement set  
 number of samplings: 1  
 sampling time: 30 min  
 sampling volume: 2,0 l/min  
 H<sub>2</sub>O mass: 4,25 g  
 absorption efficiency: 98,7 %

**O<sub>2</sub> paramagnetic (PN-EN 14789:2017):**

Range : 0 – 5 % [ ]  
           0 – 10 % [X]  
           0 – 25 % [ ]  
 Calibration gas: O<sub>2</sub> – 8,001% [ ] R/07/W  
                   O<sub>2</sub> – 18,85% [X] R/12/W  
                   N<sub>2</sub> – 99,99% [X] R/17/  
 Sampling train: 2 measurement axis  
 Sampling time/ average time: 360 min / 1 min

O <sub>2</sub> concentration measurement (paramagnetic):		HORIBA PG-350E-EU (130/1/AW/21)		
check operation		unit	result	criterion
„0” after adjustment, without sampling train		%	0,05	± 0,1 [%]*
„0” after adjustment, with sampling train	before measurement	%	0,1	± 0,5 [%]**
	after measurement	%	0,11	± 0,5 [%]**
„Span” after adjustment with sampling train /standard 18,85 % R/12/W analyzer range 25%	before measurement	%	18,9	± 0,5 [%]**
	after measurement	%	18,8	± 0,5 [%]**
	before measurement time T90	%	8,2	> 17,0 [%]
		sec.	32	< 200 sec.

**Test Report No PW/25/11/23****7. MEASUREMENT DEVICES**

Name of measuring device	X1- Apis PLUS
Type of measuring device	Isokinetic sampler S/N 0185
Certificate	Calibration No 94/54/LA/P/2023 G-73/23-46/23 65/1/T/23
Issued by	PLUM Sp. z o.o. KLEOSIN ZAP BESTWINKA LABOSERWIS Sp. z o.o. KATOWICE
Date of issue the certificate of calibration	27.02.2023 r. 23.02.2023 r. 02.03.2023 r.
Expiration date of the certificate of calibration	-

Name of measuring device	HORIBA
Type of measuring device	PG-350E-EU
Certificate	Calibration No 130/1/AW/21
Issued by	Laboserwis Sp. z o.o. Katowice
Date of issue the certificate of calibration	28.05.2021
Expiration date of the certificate of calibration	-

Name of measuring device	Sampler
Type of measuring device	SKC 224-PCMTX4, 224-44MTX
Certificate	Calibration No n.d.
Issued by	n.d.
Date of issue the certificate of calibration	n.d.
Expiration date of the certificate of calibration	-

Name of measuring device	SKC SAMPLER KIT
Type of measuring device	-
Certificate	Calibration No T271/PMT/2021 212/1/C/21
Issued by	CZACH-POMIAR Sp. z o.o. KATOWICE LABOSERWIS Sp. z o.o. KATOWICE
Date of issue the certificate of calibration	21.09.2021 01.09.2021
Expiration date of the certificate of calibration	-

**Test Report No PW/25/11/23****8. ANALYSIS RESULTS****Attachment no. 1 to the Certificate of Analysis for work order PR23C3884**

Sample: P/38/10/23

**Measurement results PCDD/Fs:**

Sample:		P/38/10/23	Final extract [µL]:	60			
		Injection volume [µL]:	4				
		Acquisition date [d.m.y h:m]:	12.11.23 19:48				
2,3,7,8-PCDD/Fs	Result [ng/sample]	Limit of Detection [ng/sample]	Limit of Quantification [ng/sample]	I-TEFs	I-TBQ		
2,3,7,8-TCDD	< 0.0042	0.0042	0.0085	1	0.0042		
1,2,3,7,8-PeCDD	< 0.0049	0.0049	0.0098	0.5	0.0025		
1,2,3,4,7,8-HxCDD	< 0.0093	0.0093	0.019	0.1	0.00093		
1,2,3,6,7,8-HxCDD	< 0.0093	0.0093	0.019	0.1	0.00093		
1,2,3,7,8,9-HxCDD	< 0.0093	0.0093	0.019	0.1	0.00093		
1,2,3,4,6,7,8-HxCDD	< 0.015	0.015	0.031	0.01	0.00015		
OCDD	< 0.029	0.029	0.058	0.001	0.000029		
2,3,7,8-TCDF	< 0.0042	0.0042	0.0084	0.1	0.00042		
1,2,3,7,8-PeCDF	< 0.0048	0.0048	0.0097	0.05	0.00024		
2,3,4,7,8-PeCDF	< 0.0048	0.0048	0.0097	0.5	0.0024		
1,2,3,4,7,8-HxCDF	< 0.0089	0.0089	0.018	0.1	0.00089		
1,2,3,6,7,8-HxCDF	< 0.0089	0.0089	0.018	0.1	0.00089		
1,2,3,7,8,9-HxCDF	< 0.0089	0.0089	0.018	0.1	0.00089		
2,3,4,6,7,8-HxCDF	< 0.0089	0.0089	0.018	0.1	0.00089		
1,2,3,4,6,7,8-HxCDF	< 0.014	0.014	0.028	0.01	0.00014		
1,2,3,4,7,8,9-HxCDF	< 0.014	0.014	0.028	0.01	0.00014		
OCDF	< 0.025	0.025	0.049	0.001	0.000025		
I-TEQ from quantified 2,3,7,8-PCDD/Fs - "Lowerbound"				0			
I-TEQ from 2,3,7,8-PCDD/Fs - „Mediumbound”				0.0083			
Maximum possible I-TEQ - "Upperbound"				0.017			
PCDDs	Result [ng/sample]	PCDFs	Result [ng/sample]				
Tetra-CDDs	< 0.093	Tetra-CDFs	< 0.16				
Penta-CDDs	< 0.069	Penta-CDFs	< 0.14				
Hexa-CDDs	< 0.093	Hexa-CDFs	< 0.14				
Hepta-CDDs	< 0.031	Hepta-CDFs	< 0.056				
OCDD	< 0.029	OCDF	< 0.025				

I-TEF according to NATO.

Limits of quantification are defined as double of the detection limits.

The limit of detection is defined as the amount of analyte producing a signal with S/N&gt;3.

The value of the detection limit is mentioned as the actual value at the acquisition date.

Measurement uncertainty is expressed as a double ( $k=2$ ) relative standard deviation (RSD%), and corresponds to 95% confidence interval.

Estimation of uncertainty of each 2,3,7,8-PCDD/F congener is 30% and total I-TEQ is 20%.

These values were ensured by analyses of certified reference material under conditions of internal reproducibility.

Results marked with "&lt;" are below limit of detection or quantification.

"Lowerbound" and "Upperbound" are levels defined in Regulation 2017/644 and EN 1948-4.

"Mediumbound" is levels defined in Regulation 2017/644.

**Test Report No PW/25/11/23****Attachment no. 1 to the Certificate of Analysis for work order PR23C3884**

Sample: P/38/10/23

**Standards recovery:**

Sample:	P/38/10/23				
Extraction standard	Recovery	Acceptable range [%]		Accept. rec. with respect to	
	[%]	Basic	Extended	basic range	extended range
PCDDs					
13C12 - 2,3,7,8-TCDD	91	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,7,8-PeCDD	60	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,4,7,8-HxCDD	75	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,6,7,8-HxCDD	83	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,4,6,7,8-HpCDD	43	40 - 130	20 - 150	YES	-
13C12 - OCDD	40	40 - 130	20 - 150	YES	-
PCDFs					
13C12 - 2,3,7,8-TCDF	84	50 - 130	30 - 150	YES	-
13C12 - 2,3,4,7,8-PeCDF	77	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,4,7,8-HxCDF	72	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,6,7,8-HxCDF	84	50 - 130	30 - 150	YES	-
13C12 - 2,3,4,6,7,8-HxCDF	54	50 - 130	30 - 150	YES	-
13C12 - 1,2,3,4,6,7,8-HpCDF	47	40 - 130	20 - 150	YES	-
13C12 - OCDF	43	40 - 130	20 - 150	YES	-
Sampling standard	Recovery	Acceptable range		Rec. in range?	
	[%]	[%]			
13C12-1,2,3,7,8-PeCDF	90	> 50		YES	
13C12-1,2,3,7,8,9-HxCDF	84	> 50		YES	
13C12-1,2,3,4,7,8,9-HpCDF	90	> 50		YES	

**Test Report No PW/25/11/23****Attachment no. 3 to the Certificate of Analysis for work order PR23C3884**

Sample: P/38/10/23

ALS SAMPLE ID: PR23C3884/ 001

Measurement results PCBs:

Sample: P/38/10/23			Final extract [µl]:	250
PCBs	Result [ng/sample]	Limit of Detection [ng/sample]	Limit of Quantification [ng/sample]	*WHO-TEFs WHO-TEQ Upperbound [ng/sample]
PCB #77	< 0.051	0.051	0.17	0.0001 0.0000051
PCB #81	< 0.042	0.042	0.14	0.0003 0.000013
PCB #126	< 0.04	0.04	0.13	0.1 0.004
PCB #169	< 0.069	0.069	0.23	0.03 0.0021
PCB #105	< 0.083	0.083	0.28	0.00003 0.0000025
PCB #114	< 0.041	0.041	0.14	0.00003 0.0000012
PCB #118	< 0.15	0.045	0.15	0.00003 0.0000045
PCB #123	< 0.074	0.074	0.25	0.00003 0.0000022
PCB #156	< 0.063	0.063	0.21	0.00003 0.0000019
PCB #157	< 0.062	0.062	0.21	0.00003 0.0000019
PCB #167	< 0.051	0.051	0.17	0.00003 0.0000015
PCB #170	< 0.077	0.077	0.26	- 0
PCB #180	< 0.068	0.068	0.23	- 0
PCB #189	< 0.057	0.057	0.19	0.00003 0.0000017
WHO-TEQ from quantified PCBs -"Lowerbound"				
WHO-TEQ from PCBs -"Mediumbound"				
Maximum possible WHO-TEQ -"Upperbound"				
PCBs	Result [ng/sample]	Limit of Detection [ng/sample]	Limit of Quantification [ng/sample]	Eindicator PCB Lowerbound [ng/sample] Eindicator PCB Upperbound [ng/sample]
PCB #28	< 1.2	0.066	1.2	0 1.2
PCB #52	< 0.26	0.063	0.26	0 0.26
PCB #101	< 0.32	0.054	0.32	0 0.32
PCB #118	< 0.15	0.045	0.15	0 0.15
PCB #138	< 0.21	0.064	0.21	0 0.21
PCB #153	< 0.28	0.055	0.28	0 0.28
PCB #180	< 0.068	0.068	0.23	0 0.068
Eindicator PCB6 -"Lowerbound"				
Maximal possible Eindicator PCB6 -"Upperbound"				
Eindicator PCB7 -"Lowerbound"				
Maximal possible Eindicator PCB7 -"Upperbound"				

\*WHO 2005 TEF according to Van den Berg et al: Toxicological Sciences Advance Acces, 7 July 2006

Limits of quantification are defined on the base of blank level.

The limit of detection is defined as the amount of analyte producing a signal with S/N&gt;3.

The value of the detection limit is mentioned as the actual value at the acquisition date.

Measurement uncertainty is expressed as a double (k=2) relative standard deviation (RSD%), and corresponds to 95% confidence interval.

Estimation of uncertainty of each PCB congener is 30%, total WHO-TEQ and PCB6/PCB7 is 20%. These values were ensured by analyses of certified reference material under conditions of internal reproducibility.

Results marked "&lt;" are lower than the limit of detection or quantification.

"Lowerbound" and "Upperbound" are levels defined in Regulation 2017/644 and EN 1948-4.

"Mediumbound" is level defined in Regulation 2017/644.

**Test Report No PW/25/11/23*****Attachment no. 3 to the Certificate of Analysis for work order PR23C3884***


Sample:	P/38/10/23	Final extract [ $\mu$ L]:	250
		Injection volume [ $\mu$ L]:	4
		Acquisition date [d.m.y h:m]:	12.11.23 21:12
Sampling standard	Recovery [%]	Acceptable range [%]	Rec. in range?
13C12-2,3,4,4'-tetraCB (60)	70	> 50	YES
13C12-3,3',4,5,5'-pentaCB (127)	93	> 50	YES
13C12-2,3,3',4,5,5'-hexaCB (159)	73	> 50	YES

**Test Report No PW/25/11/23****9. CERTIFICATE OF ACCREDITATION****POLSKIE CENTRUM AKREDYTACJI**  
**POLISH CENTRE FOR ACCREDITATION**Sygnatariusz EA MLA  
EA MLA Signatory**CERTYFIKAT AKREDYTACJI**  
**LABORATORIUM BADAWCZEGO**  
**ACCREDITATION CERTIFICATE OF TESTING LABORATORY**  
**Nr AB 994**

Potwierdza się, że: / This is to confirm that:

**„PROFTECH” Sp. z o.o.**  
ul. Kurta Aldera 44, 41-506 Chorzówspełnia wymagania normy PN-EN ISO/IEC 17025:2018-02  
meets requirements of the PN-EN ISO/IEC 17025:2018-02 standardAkredytowana działalność jest określona w Zakresie Akredytacji Nr AB 994  
Accredited activity is defined in the Scope of Accreditation No AB 994Akredytacja pozostaje w mocy pod warunkiem przestrzegania  
wymagań jednostki akredytującej określonych w kontrakcie Nr AB 994  
This accreditation remains in force provided the Laboratory observes  
the requirements of Accreditation Body defined in the Contract No AB 994Akredytacji udzielono dnia 30.01.2009 r.  
Accreditation was granted on 30.01.2009DYREKTOR  
POLSKIEGO CENTRUM AKREDYTACJI  
LUCYNA OLBORSKA

Warszawa, dnia 9 grudnia 2019 roku

**Test Report No PW/25/11/23****10. MEASUREMENT PLANE SCHEME**

Approved by  
  
DYREKTOR  
Name and Signature  
inż. Dariusz Guja

**END OF REPORT**

